

कील, चौकोर, रेडियस — विशिष्टि

(पहला पुनरीक्षण)

Nail, Square, Radius — Specification

(*First Revision*)

ICS 11.040.40

© BIS 2023



भारतीय मानक ब्यूरो
BUREAU OF INDIAN STANDARDS
मानक भवन, 9 बहादुर शाह ज़फर मार्ग, नई दिल्ली - 110002
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI - 110002
www.bis.gov.in www.standardsbis.in

FOREWORD

This Indian Standard (First Revision) was adopted by the Bureau of Indian Standards, after the draft finalized by the Orthopaedic Instruments, Implants and Accessories Sectional Committee had been approved by the Medical Equipment and Hospital Planning Division Council.

This standard was originally published in 1983. The first revision of this standard has been brought out to align it with the recent developments and to bring the standard in line with the latest style and format of Indian Standards.

For the purpose of deciding whether a particular requirement of this standard is complied with the final value, observed or calculated, expressing the result of a test or analysis shall be rounded off in accordance with IS 2: 2022 'Rules for rounding off numerical values (*second revision*)'. The number of significant places retained in the rounded off value should be same as that of the specified value in this standard.

Indian Standard

NAIL, SQUARE, RADIUS — SPECIFICATION

(First Revision)

1 SCOPE

This standard prescribes dimensional and other requirements of square nails used for implantation of radius bone in orthopaedic surgery.

2 REFERENCES

The standards given below contain provisions which, through reference in this text, constitute provisions of this standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent edition of these standards.

<i>IS No.</i>	<i>Title</i>
IS/ISO 5832 (Part 1) : 2016	Implants for surgery — Metallic materials: Part 1 Wrought stainless steel
IS 7531 : 1990	Surgical instruments — Corrosion resistance of stainless-steel surgical instruments — Methods of tests (<i>first revision</i>)

3 MATERIAL

The composition of stainless steel shall be as per Table 1 given in 4.2 of IS/ISO 5832 (Part 1).

4 SHAPE AND DIMENSIONS

4.1 The dimensions shall be as shown in Fig. 1.

4.2 The diameter of square nails may be calculated by multiplying by $\sqrt{2}$ when required for the purpose of indentation.

5 WORKMANSHIP AND FINISH

5.1 One end of the nail shall be as shown in Fig. 1 and the other end threaded.

5.2 The thread shall be neat and clean.

5.3 The nail shall be free from cracks, draw marks, burrs, pits, surface contamination and foreign particles.

5.4 The nail shall be electro-polished and passivated.

6 CORROSION RESISTANCE TEST

The product shall satisfy the requirements when tested in accordance with IS 7531.

7 MARKING


7.1 Each component shall be marked with manufacturer's name, initials or registered trademark.

7.2 BIS Certification Marking


The product(s) conforming to the requirements of this standard may be certified as per the conformity assessment schemes under the provisions of the *Bureau of Indian Standards Act, 2016* and the Rules and Regulations framed there under, and the product(s) may be marked with the Standard Mark.

8 PACKING

The tip of each nail may be covered with strippable plastic coating. The nail may then be packed individually in polyethylene bags. A number of such pieces may be packed in cartons as agreed to between the purchaser and the supplier.



CALCULATION FOR DIAMETER
OF SQUARE NAIL



$D = 2.0$ to 4.0 in steps of 0.5 .
 $L = 150$ to 275 in steps of 25 .
All dimensions in millimetres.

FIG. 1 NAIL, SQUARE, RADIUS

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 2016* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Head (Publication & Sales), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the website- www.bis.gov.in or www.standardsbis.in.

This Indian Standard has been developed from Doc No.: MHD 02 (20843).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002
Telephones: 2323 0131, 2323 3375, 2323 9402

Website: www.bis.gov.in

Regional Offices:

Central	: 601/A, Konnectus Tower -1, 6 th Floor, DMRC Building, Bhavbhuti Marg, New Delhi 110002	{ 2323 7617
Eastern	: 8 th Floor, Plot No 7/7 & 7/8, CP Block, Sector V, Salt Lake, Kolkata, West Bengal 700091	{ 2367 0012 2320 9474
Northern	: Plot No. 4-A, Sector 27-B, Madhya Marg, Chandigarh 160019	{ 265 9930
Southern	: C.I.T. Campus, IV Cross Road, Taramani, Chennai 600113	{ 2254 1442 2254 1216
Western	: Plot No. E-9, Road No.-8, MIDC, Andheri (East), Mumbai 400093	{ 2821 8093

Branches : AHMEDABAD. BENGALURU. BHOPAL. BHUBANESHWAR. CHANDIGARH. CHENNAI.
COIMBATORE. DEHRADUN. DELHI. FARIDABAD. GHAZIABAD. GUWAHATI.
HIMACHAL PRADESH. HUBLI. HYDERABAD. JAIPUR. JAMMU & KASHMIR.
JAMSHEDPUR. KOCHI. KOLKATA. LUCKNOW. MADURAI. MUMBAI. NAGPUR.
NOIDA. PANIPAT. PATNA. PUNE. RAIPUR. RAJKOT. SURAT. VISAKHAPATNAM.